

**CERTIFICATE OF ANALYSIS**

Product: OptiLube Sterile Lubricating Jelly
Description: 5g Sachet
Product reference: 1120
Lot number: 090121
Manufacture date: 21/04/2021
Expiry date: 01/04/2026
Sterilisation method: Gamma Irradiation

Test	Specification	Results
Appearance	Gel/Jelly	Pass
Colour	Colourless to pale yellow	Pass
Solubility in water	Soluble	Pass
Packaging	Seal	Pass
pH	5.0 - 7.0	6.13
Viscosity	13000 - 18000 cP	16640 cP

This testing has been carried out by a third-party company.

Optimum Medical hereby certifies that the above listed product has been irradiated by Gamma Rays and meets dose requirements. Actual absorbed dose has been measured through a dosimetric system with acceptable precision and accuracy.

Cleo Atkins

QA RA Associate

28/01/2022

Cert 4 Issue 3

